



OFFICE OF THE DEPARTMENT OF DEFENSE COORDINATOR
FOR DRUG ENFORCEMENT POLICY AND SUPPORT

1510 DEFENSE PENTAGON
WASHINGTON DC 20301-1510



29 JUN 1995

MEMORANDUM FOR UNDER SECRETARY OF THE NAVY (ATTN: CAPT WEISBERG)
ASSISTANT SECRETARY OF THE ARMY (IL&E)
DEPUTY ASSISTANT SECRETARY OF THE AIR FORCE (RA&CDP)

SUBJECT: DoD Laboratory Certification for Reporting Drug
Screening by Immunochemistry Analysis on Olympus AU800
Analyzers

As a result of First Article Testing of the Roche reagents conducted on the Olympus AU800, this memorandum is to clarify operational issues relative to my memorandum dated June 19, 1995 on the subject topic. First, for the laboratory certification, the Olympus AU800 should be calibrated using a four point calibration. Second, delete the requirements for specimens at 50% of the calibrator concentration and in section 1.b for the negative control to have a coefficient of variation of 7.5%. Third, add to Paragraph 2, Carryover Check, the following:

b) Two batch runs containing specimens in the following sequence:

- four 500% calibration control samples
- four 75% calibration control samples
- two negative urine samples
- four 500% calibration control samples
- four negative urine samples

My point of contact for this matter is Captain John F. Jemionek, MSC, USN who may be reached at (703) 693-1917.

Brian E. Sheridan
Deputy Assistant Secretary for
Drug Enforcement Policy and Support

CF:

FTDTL Fort Meade	ATTN: COL Armitage ✓
FTDTL Tripler AMC	ATTN: COL Jacobs ✓
NDSL S.D.	ATTN: CAPT Christophers ✓
NDSL JAX	ATTN: CDR Past ✓
NDSL GLakes	ATTN: CDR Thomas ✓
AL/AOT Brooks AFB	ATTN: LTC Mehm ✓
AFIP, WRAMC	ATTN: COL Smith ✓
OTSG, ARMY	ATTN: COL O'Brien ✓
BUMED, MED-244	ATTN: CAPT Hughes ✓
USADAO	ATTN: Mr. Cunningham

